

Repair coupon AirOne / AirOne Flex

Please enclose with the package.

Service department for Germany and Austria:

Aesculap Schermaschinen GmbH
Fröhliche-Mann-Straße 15
98528 Suhl
Service department Hippomed
service@aesculap-schermaschinen.de
Tel.: 03681 4982-99 Fax: 03681 4982-34

Information service department outside Germany / Austria:

Tel: +49 (0) 8086 / 933-327
E-mail: inhalatoren@kerbl.com

Warranty claims only valid with warranty card or proof of purchase

| |
|---------------------------------------|
| Dealer's stamp |
| |
| VAT No.: _____ |
| Contact: _____ |
| Telephone number (+ extension): _____ |
| E-Mail or Fax*: _____ |

| |
|---|
| Customer address |
| |
| Surname, first name _____ |
| Address line 1 _____ |
| Address line 2, Postcode _____ |
| Telephone number for any questions: _____ |
| E-Mail or Fax*: _____ |

Return to*: Dealer's address

Customer address

Customer requests*:

Warranty (enclose proof of purchase and/or completed warranty card)

Important information:

Kerbl item no.: _____

Device name: _____

Serialnumber: _____

IMPORTANT: No work may be carried out without a warranty card or invoice.

Repair (without estimate) by €

Estimate/repair flat fee

We would like to point out that there is a fee of € 50.00 excl. VAT for the preparation of a cost estimate or the examination of the device. This amount will only be charged if no repairs are to be carried out.

| |
|---|
| fault description*: |
| <input type="radio"/> device nebulises too weakly |
| <input type="radio"/> does not nebulise at all |
| <input type="radio"/> keyboard on / off does not work |
| <input type="radio"/> the water level indicator lights up |
| <input type="radio"/> the water temperature display lights up continuously, although there is sufficient contact liquid |
| <input type="radio"/> other errors |